



## **National Accreditation Board for Testing and Calibration Laboratories**

(A Constituent Board of Quality Council of India)

NABL / C-0856 22.07.2023

MANOJ V DHUMAL

S G SYSTEM

FLAT NO. 30, NEW RAJGAD ASSOCIATES, BALAJI NAGAR, DHANKAWADI PUNE,MAHARASHTRA-411043

Mobile: 8007940295

E-mail: sgsysteml 1@gmail.com

Subject: Renewal of Accreditation

Dear Sir,

Refer letter no. NABL / C-0856 13.07.2023, NABL has reviewed reply submitted by laboratory and decided following:

NABL has reviewed reply submitted by laboratory & and decided to

- Inclusion of Pressure Indicating Devices {(-) 0.85 bar to 0} with CMC 0.007 bar in view of revised CMC calculations submitted by the laboratory.
- Inclusion of Electronic Height Gauge (L.C.: 0.0001 mm & coarser) in view of the clarifications and revised scope submitted by the laboratory.

The accreditation certificate no.CC-3639 with issue dates 11/07/2023 and valid till 10/07/2025 is under preparation stage.

You have to submit the Desktop surveillance application two months prior to the expiry of first year of accreditation cycle to maintain continuity in Accreditation cycle. Kindly Treat this letter reminder for Desktop Surveillance.

Yours Sincerely, **Gopinath Tanneeru** gopinath@nabl.qcin.org

> NABL House, Plot 45, Sector 44, Gurugram 122 003, Haryana, India Tel. No.: +91-124-4679700 (30 lines) \* Fax: +91-124-4679799 \* Website: www.nabl-india.org