



MIKRONIX CALIBRATION CENTRE  
AURANGABAD 431 210

CALIBRATION CERTIFICATE OF  
ULTRASONIC THICKNESS GAUGE

Calibration Date	Issued On	Certificate No.	Next Due Date	Page	No. of Pages
13/08/2022	13/08/2022	INST/22/9162	13/08/2023 (cust. decided)	1	2

Name of Customer & Address : M/s. CHARTERED TECHNO SERVICES  
'D' WING 'A' BLOCK, 1ST FLOOR, BHARAT BAZAR, API CORNER, CHIKALTHANA, AURANGABAD - 431 210.,

Date of Receipt of Material : 13/08/2022 (Challan No. : CTS/R.Extra/53/2022 dtd. 12/08/2022)

Work Order No. : 334

Unit Under Calibration : ULTRASONIC THICKNESS GAUGE

Make : Megasonic

Size : --

Sr. No. : M820

Identification No. : DT1-C

Lab ID No. : 0334/001

Condition of Item : Ok

Environmental Conditions : Temperature : 20.2°C Humidity : 50±10% Rh

Reference Document No. :

Calibration Procedure No. :

Major Equipments Used for Calibration : Used standards are traceable to National / International Standards through Direct / NABL Accredited Calibration Laboratory.

Description	Identification No.	Calibration Certificate No	Calibration Date	Valid up to	Traceability With
Steel Slip Gauge Set M122/1 Gr. 'K'	86344	P/01/200039	19/08/2020	19/08/2022	CC-2082

Uncertainty of Measurement : ± 0.1 mm

(At 95% confidence level with coverage factor K=2)

Note:1) Calibration Results are given on Page No.2 onwards.

Note : The Calibration results reported in this certificate are valid at the time of and the stated condition of environment.  
This certificate shall not be reproduced except in full without written permission from Proprietor,  
Mikronix Calibration Centre, Aurangabad.

CONTD. PAGE 2

  
CALIBRATED BY

  
ASS. TECHNICAL MANAGER

MCC/FRM-A4/33

MIKRONIX CALIBRATION CENTRE

OFFICE & LABORATORY : PLOT NO. 12, CIDCO SERVICE INDUSTRIAL ZONE,  
OPP. PROZONE MALL, AURANGABAD - 431 006. (M.S.) INDIA.  
☎ : (0240) 2487530, Cell : 8605010301.  
E-mail : mikronix\_cal@yahoo.com • Website : www.mikronixassociates.com