



National Accreditation Board for  
Testing and Calibration Laboratories

**CERTIFICATE OF ACCREDITATION**

**PRECISION INSTRUMENTATION AND SERVICES PVT.LTD**

has been assessed and accredited in accordance with the standard

**ISO/IEC 17025:2017**

**"General Requirements for the Competence of Testing &  
Calibration Laboratories"**

for its facilities at

21, SHREERAM BHAVAN, F-9, SHREERAM KUNJ, TAKLI ROAD, DWARKA, NASHIK, MAHARASHTRA,  
INDIA

in the field of

**CALIBRATION**

Certificate Number: CC-2098

Issue Date: 29/12/2021

Valid Until: 28/12/2023

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.  
(To see the scope of accreditation of this laboratory, you may also visit NABL website [www.nabl-india.org](http://www.nabl-india.org))

Name of Legal Identity : PRECISION INSTRUMENTATION AND SERVICES PRIVATE LIMITED

Signed for and on behalf of NABL



N. Venkateswaran  
Chief Executive Officer



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:- 1 of 1
<b>Bush Electromech and Engineering Pvt. Ltd.</b>	Date of Receipt	:- 6-Oct-22
K-89, MIDC, Waluj, Aurangabad 431 136.	Service Request No.	:- A2022/22
Ambient Temp. :- 24.5 °C	Certificate No.	:- A2022/10/065
Relative Humidity :- 52 % RH	Date of Calibration	:- 6-Oct-22
Location of calibration :- On Site	Next Calibration Due On	:- 6-Oct-23
Condition of Item :- OK	Calibration method No.	:- PISPL/ET/WI/03

<b>2. Details of unit under Calibration (UUC):-</b>			
Name :- AC High Voltage Tester	Range	:- 0 to 5 KVAC	
I.D No. :- 101101	Resolution	:- ---	
Make :- Agronics	Specified Accuracy	:- ± 2 %	
	Location	:- ---	

<b>3.Details of Equipment Used for Calibration</b>			
Name :- High Voltage Probe	Digital Multimeter		
Make/I.D No. :- Fluke/ PIS/HV/01	CIE/PISPL/ABAD/DMM/02		
Certificate No. :- TSC/20-21/13518-2	E2020/12/298		
Certified By :- Transcal	PISPL		
Calibration Validity :- 18-Mar-22	21-Dec-21		

<b>4.Calibration Results</b> :-				
Calibration Points KVAC	UUC Reading KVAC	Standard Reading KVAC	Error in KVAC	Error in %
1	1	0.996	0.004	0.40
2	2	1.994	0.006	0.30
3	3	2.953	0.047	1.57
4	4	3.981	0.019	0.48
5	5	4.977	0.023	0.46


**5.Accuracy in Measurement** :- ± 1.57 %

**6.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

**Note:**

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration.
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
Rahul  
Calibration Engineer  
Mr.Rahul Pawar



Approved By  
ceio  
Technical Manager  
Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:- 1 of 1
<b>Bush Electromech and Engineering Pvt. Ltd.</b>	Date of Receipt	:- 6-Oct-22
K-89, MIDC, Waluj, Aurangabad 431 136.	Service Request No.	:- A2022/22
Ambient Temp. :- 24.2 °C	Certificate No.	:- A2022/10/066
Relative Humidity :- 52 % RH	Date of Calibration	:- 6-Oct-22
Location of calibration :- On Site	Next Calibration Due On	:- 6-Oct-23
Condition of Item :- OK	Calibration method No.	:- PISPL/ET/WI/03

<b>2. Details of unit under Calibration (UUC):-</b>		
Name :- AC High Voltage Tester	Range	:- 0 to 50 KVAC
I.D No. :- 8637/174(P)/09-10	Resolution	:- 2 KVAC
Make :- Rectifier & Electronics	Specified Accuracy	:- ± 2 %
	Location	:- ---

<b>3.Details of Equipment Used for Calibration</b>		
Name :- High Voltage Probe	Digital Multimeter	
Make/I.D No. :- Fluke/ PIS/HV/01	Rish/PIS/DMM/02	
Certificate No. :- TSC/21-22/18060-1	TSC/21-22/18060-1	
Certified By :- Transcal	Transcal	
Calibration Validity :- 14-Mar-23	14-Mar-23	


<b>4.Calibration Results</b> :-				
Calibration Points KVAC	UUC Reading KVAC	Standard Reading KVAC	Error in KVAC	Error in %
5	5	4.995	0.005	0.10
10	10	9.982	0.018	0.18
15	15	14.977	0.023	0.15
20	20	19.966	0.034	0.17
28	28	27.962	0.038	0.14

**5.Accuracy in Measurement** :- ± 0.18%


**6.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

**Note:**

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
  
Calibration Engineer  
Mr.Rahul Pawar



Approved By  
  
Technical Manager  
Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>Bush Electromech and Engineering Pvt. Ltd.</b> K-89, MIDC, Waluj, Aurangabad 431 136. Ambient Temp. :- 24.2 °C Relative Humidity :- 52 % RH Location of calibration :- On Site Condition of Item :- OK	Page No. :- 1 of 1 Date of Receipt :- 6-Oct-22 Service Request No. :- A2022/22 Certificate No. :- A2022/10/067 Date of Calibration :- 6-Oct-22 Next Calibration Due On :- 6-Oct-23 Calibration method No. :- PISPL/ET/WI/02
--	---

<b>2. Details of unit under Calibration (UUC):-</b>			
Name :- AC High Voltage Tester (Ammeter)	Range :- 0 to 100 mAAC	Resolution :- 4 mAAC	Specified Accuracy :- ± 2 %
I.D No. :- 8637/174(P)/09-10	Location :- ---		
Make :- Rectifier & Electronics			

<b>3.Details of Equipment Used for Calibration</b>	
Name :- Multifunction Calibration System	
Make/I.D No. :- Fluke / PIS/UNICAL/03	
Certificate No. :- M-220829-20-2	
Certified By :- GODREJ	
Calibration Validity :- 29-Aug-23	

<b>4.Calibration Results</b> :-				
Calibration Points mAAC	UUC Reading mAAC	Standard Reading mAAC	Error in mAAC	Error in %
20	20	19.955	0.045	0.23
40	40	39.927	0.073	0.18
60	60	59.871	0.129	0.21
80	80	79.743	0.257	0.32
100	100	99.597	0.403	0.40

**5.Accuracy in Measurement** :- ± 0.40 %

**6.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

- Note:**
- 1) UUC stands for Unit Under Calibration.
  - 2) This certificate refers only to the particular item submitted for calibration
  - 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
  - 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
Rahul

Calibration Engineer  
Mr.Rahul Pawar



Approved By  
Vinita

Technical Manager  
Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:-	1 of 1
<b>Bush Electromech and Engineering Pvt. Ltd.</b>	Date of Receipt	:-	6-Oct-22
K-89, MIDC, Waluj, Aurangabad 431 136.	Service Request No.	:-	A2022/22
Ambient Temp. :- 24.2 °C	Certificate No.	:-	A2022/10/068
Relative Humidity :- 52 % RH	Date of Calibration	:-	6-Oct-22
Location of calibration :- On Site	Next Calibration Due On	:-	6-Oct-23
Condition of Item :- OK	Calibration method No.	:-	PISPL/ET/WI/03

### 2. Details of unit under Calibration (UUC):-

Name :- AC High Voltage Tester	Range :-	0 to 100 KVAC
I.D No. :- NTPL/055/17-18	Resolution :-	0.1 KVAC
Make :- ntpl	Specified Accuracy :-	± 2 %
	Location :-	---

### 3.Details of Equipment Used for Calibration

Name :- High Voltage Probe	Digital Multimeter
Make/I.D No. :- Fluke/ PIS/HV/01	Rish/PIS/DMM/02
Certificate No. :- TSC/21-22/18060-1	TSC/21-22/18060-1
Certified By :- Transcal	Transcal
Calibration Validity :- 14-Mar-23	14-Mar-23

### 4.Calibration Results :-

Calibration Points KVAC	UUC Reading KVAC	Standard Reading KVAC	Error in KVAC	Error in %
5	5	4.988	0.012	0.24
10	10	9.984	0.016	0.16
15	15	14.962	0.038	0.25
20	20	19.951	0.049	0.24
28	28	27.942	0.058	0.21

**5.Accuracy in Measurement** :- ± 0.25 %

**6.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

#### Note:

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
Rahul  
Calibration Engineer  
Mr.Rahul Pawar



Approved By  
Cele  
Technical Manager  
Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:-	1 of 1
<b>Bush Electromech and Engineering Pvt. Ltd.</b>	Date of Receipt	:-	6-Oct-22
K-89, MIDC, Waluj, Aurangabad 431 136.	Service Request No.	:-	A2022/22
Ambient Temp. :- 24.2 °C	Certificate No.	:-	A2022/10/069
Relative Humidity :- 52 % RH	Date of Calibration	:-	6-Oct-22
Location of calibration :- On Site	Next Calibration Due On	:-	6-Oct-23
Condition of Item :- OK	Calibration method No.	:-	PISPL/ET/WI/02

### 2. Details of unit under Calibration (UUC):-

Name :- AC High Voltage Tester (Ammeter)	Range	:-	0 to 100 mAAC
I.D No. :- NTPL/055/17-18	Resolution	:-	0.1 mAAC
Make :- ntpl	Specified Accuracy	:-	± 2 %
	Location	:-	---

### 3.Details of Equipment Used for Calibration

Name :- Multifunction Calibration System
Make/I.D No. :- Fluke / PIS/UNICAL/03
Certificate No. :- M-220829-20-2
Certified By :- GODREJ
Calibration Validity :- 29-Aug-23

### 4.Calibration Results :-

Calibration Points mAAC	UUC Reading mAAC	Standard Reading mAAC	Error in mAAC	Error in %
20	20.0	19.967	0.033	0.17
40	40.0	39.920	0.080	0.20
60	60.0	59.864	0.136	0.23
80	80.0	79.755	0.245	0.31
100	100.0	99.637	0.363	0.36

**5.Accuracy in Measurement** :- ± 0.36 %

**6.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

#### Note:

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
*Rahul*

Calibration Engineer  
Mr.Rahul Pawar



Approved By  
*ceee*

Technical Manager  
Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:- 1 of 1
<b>Bush Electromech and Engineering Pvt. Ltd.</b>	Date of Receipt	:- 6-Oct-22
K-89, MIDC, Waluj, Aurangabad 431 136.	Service Request No.	:- A2022/22
Ambient Temp. :- 24.2 <sup>o</sup> C	Certificate No.	:- A2022/10/070
Relative Humidity :- 52 % RH	Date of Calibration	:- 6-Oct-22
Location of calibration :- On Site	Next Calibration Due On	:- 6-Oct-22
Condition of Item :- OK	Calibration method No.	:- PISPL/ET/WI/03

### 2. Details of unit under Calibration (UUC):-

Name :- AC High Voltage Tester	Range	:- 0 to 2 KVAC
I.D No. :- 1101101	Resolution	:- ---
Make :- Agronics	Specified Accuracy	:- $\pm 2\%$
	Location	:- ---

### 3.Details of Equipment Used for Calibration

Name :- High Voltage Probe	Digital Multimeter
Make/I.D No. :- Fluke/ PIS/HV/01	Rish/PIS/DMM/02
Certificate No. :- TSC/21-22/18060-1	TSC/21-22/18060-1
Certified By :- Transcal	Transcal
Calibration Validity :- 14-Mar-23	14-Mar-23

### 4.Calibration Results :-

Calibration Points KVAC	UUC Reading KVAC	Standard Reading KVAC	Error in KVAC	Error in %
1	1	0.995	0.005	0.50
2	2	1.994	0.006	0.30

**5.Accuracy in Measurement** :-  $\pm 0.50\%$

**6.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

#### Note:

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By

Calibration Engineer  
Mr.Rahul Pawar



Approved By

Technical Manager  
Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>Bush Electromech and Engineering Pvt. Ltd.</b> K-89, MIDC, Waluj, Aurangabad 431 136.	Page No. :- 1 of 1
Ambient Temp. :- 24.2 °C	Date of Receipt :- 6-Oct-22
Relative Humidity :- 52 % RH	Service Request No. :- A2022/22
Location of calibration :- On Site	Certificate No. :- A2022/10/071
Condition of Item :- OK	Date of Calibration :- 6-Oct-22
	Next Calibration Due On :- 6-Oct-22
	Calibration method No. :- PISPL/ET/WI/03

### 2. Details of unit under Calibration (UUC):-

Name :- Insulation Tester	Range :- 0 to 2.5 KVAC
I.D No. :- 363167	Resolution :- ---
Make :- Meco	Specified Accuracy :- $\pm 2\%$

### 3.Details of Equipment Used for Calibration

Name :- High Voltage Probe	Digital Multimeter
Make/I.D No. :- Fluke/ PIS/HV/01	Rish/PIS/DMM/02
Certificate No. :- TSC/21-22/18060-1	TSC/21-22/18060-1
Certified By :- Transcal	Transcal
Calibration Validity :- 14-Mar-23	14-Mar-23

### 4.Calibration Results :-


Calibration Points KVAC	UUC Reading KVAC	Standard Reading KVAC	Error in KVAC	Error in %
1	1	0.998	0.002	0.20
2.5	2.5	2.492	0.008	0.32

5.Accuracy in Measurement :-  $\pm 0.32\%$

6.Remark :- The Unit Under Calibration Found to Perform Satisfactorily.

#### Note:

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
  
Calibration Engineer



Approved By  
  
Technical Manager








# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>Bush Electromech and Engineering Pvt. Ltd.</b> K-89, MIDC, Waluj, Aurangabad 431 136.		Page No.	:- 1 of 1		
Ambient Temp.	:- 24.2 °C	Date of Receipt	:- 6-Oct-22		
Relative Humidity	:- 52 % RH	Service Request No.	:- A2022/22		
Location of calibration	:- On Site	Certificate No.	:- A2022/10/072		
Condition of Item	:- OK	Date of Calibration	:- 6-Oct-22		
		Next Calibration Due On	:- 6-Oct-23		
		Calibration method No.	:- PISPL/ET/WI/04		
<b>2. Details of unit under Calibration (UUC):-</b>					
Name	:- Contact Resistance Meter	Range	:- 40 $\mu\Omega$ to 2000 $\mu\Omega$		
I.D No.	:- 5509.00.W005	Resolution	:- ---		
Make	:- Scope	Specified Accuracy	:- $\pm 2\%$		
		Location	:- ---		
<b>3.Details of Equipment Used for Calibration</b>					
Name	:- 4 Wire low resistance				
Make/I.D No.	:- Zeal/PIS/LR/01				
Certificate No.	:- 10/161 (2022-2023)				
Certified By	:- ZEAL				
Calibration Validity	:- 27-Oct-23				
<b>4.Calibration Results</b> :-					
Calibration Points $\mu\Omega$	UUC Reading $\mu\Omega$	Standard Reading $\mu\Omega$	Error in $\mu\Omega$	Error in %	
10	10	9.98	0.02	0.20	
50	50	49.87	0.13	0.26	
100	100	99.83	0.17	0.17	
<b>5.Accuracy in Measurement</b>		:- $\pm 0.20\%$			
<b>6.Remark</b>		:- The Unit Under Calibration Found to Perform Satisfactorily.			
<b>Note:</b>					
1) UUC stands for Unit Under Calibration.					
2) This certificate refers only to the particular item submitted for calibration					
3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".					
4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.					
Calibrated By 		Approved By 			
Calibration Engineer Mr.Rahul Pawar				Technical Manager Mrs.Vinita Ghogare	



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>Bush Electromech and Engineering Pvt. Ltd.</b> K-89, MIDC, Waluj, Aurangabad 431 136.	Page No. :- 1 of 2 Date of Receipt :- 6-Oct-22 Service Request No. :- A2022/22 Certificate No. :- A2022/10/073 Date of Calibration :- 6-Oct-22 Next Calibration Due On :- 6-Oct-22 Calibration method No. :- PISPL/ET/WI/02,03,04 Location of calibration :- On Site Condition of Item :- OK
---	--

### 2.Environmental Conditions

Ambient Temp. :- 23.8 °C  
Relative Humidity :- 56 % RH

### 3.Description of Item

Name :- Digital Clamp Meter  
ID.No./Sr.no :- DCM-01  
Make/Model No. :- Mecco/DT 2250 AUTO  
Range :- Multirange

### 4.Detail of equipment used for calibration

Name :- Multifunction Calibration System with Current Coil  
Make/I.D No. :- Fluke/PIS/UNICAL/03 Fluke/PIS/UNICAL/03  
Certificate No. :- M-220829-20-2 NEC/22-23/SP/065-1  
Certified By :- GODREJ NEC  
Calibration Validity :- 29-Aug-23 22-Sep-24

### 5.Calibration Results ELECTRO TECHNICAL ELECTRICAL EQUIPMENT - (Source)

Parameter	Range	Standard Reading	UUC Reading	Error	Error in %
DC Voltage	4 V	0.50000	0.497	-0.00300	-0.600
		2.00000	1.997	-0.00300	-0.150
		4.00000	3.985	-0.01500	-0.375
	40 V	5.0000	4.97	-0.0300	-0.600
		20.0000	19.92	-0.0800	-0.400
		40.0000	39.66	-0.3400	-0.850
	400 V	50.000	50.4	0.400	0.800
		200.000	200.2	0.200	0.100
		400.00	400.1	0.10	0.025
	1000V	400.00	400	0.00	0.000
		700.00	702	2.00	0.286
		999.00	1000	1.00	0.100
AC Voltage @50Hz	4 V	0.50000	0.497	-0.00300	-0.600
		2.00000	1.994	-0.00600	-0.300
		4.00000	3.985	-0.01500	-0.375
	40 V	5.0000	4.99	-0.0100	-0.200
		20.0000	19.98	-0.0200	-0.100
		40.0000	39.67	-0.3300	-0.825
	400 V	50.000	50.2	0.200	0.400
		200.000	200.5	0.500	0.250
		400.00	400.4	0.40	0.100
	750 V	500.00	501	1.00	0.200
		600.00	601	1.00	0.167
		750.00	747	-3.00	-0.400

Calibrated By  
Rahul

Calibration Engineer



Approved By

Ull

Technical Manager



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

Page No. :- 2 of 2  
Certificate No. :- A2022/10/073  
Date of Calibration :- 6-Oct-22

Parameter	Range	Standard Reading	UUC Reading	Error	Error in %
AC Current @50Hz	40 A	4.0000	4.11	0.1100	2.750
		20.0000	20.14	0.1400	0.700
		40.0000	40.18	0.1800	0.450
	400A	50.0000	50.1	0.1000	0.200
		200.000	200.2	0.200	0.100
		400.00	401.2	1.20	0.300
Resistance	400 Ω	40.0000	40.10	0.1000	0.500
		200.000	200.1	0.100	0.500
		400.00	399.9	-0.10	-0.500
	4KΩ	0.50000	0.497	-0.00300	-0.015
		4.00000	3.998	-0.00200	-0.010
	400KΩ	40.0000	39.95	-0.0500	-0.250
		400.00	400.1	0.10	0.500
	4MΩ	0.50000	0.500	0.00000	0.000
		4.00000	4.007	0.00700	0.035
	40MΩ	10.0000	10.02	0.0200	0.100
		40.0000	40.01	0.0100	0.050

### Note:

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt.Ltd, Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
*Rahul*  
Calibration Engineer  
Mr.Rahul Pawar



Approved By  
*Vinita*  
Technical Manager  
Mrs.Vinita Ghogare

PF-31/0



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>Bush Electromech and Engineering Pvt. Ltd.</b> K-89, MIDC, Waluj, Aurangabad 431 136.	Page No. :- 1 of 1
Ambient Temp. :- 24.2 °C	Date of Receipt :- 6-Oct-22
Relative Humidity :- 52 % RH	Service Request No. :- A2022/22
Location of calibration :- On Site	Certificate No. :- A2022/10/074
Condition of Item :- OK	Date of Calibration :- 6-Oct-22
	Next Calibration Due On :- 6-Oct-23
	Calibration method No. :- PISPL/ET/WI/03

<b>2. Details of unit under Calibration (UUC):-</b>			
Name :- Primary Injection Kit	Range :- 0 to 270 VAC		
Sr. No. :- 2016-40	Resolution :- ---		
Make :- Smruti	Specified Accuracy :- ± 2 %		
	Location :- ---		

<b>3.Details of Equipment Used for Calibration</b>	
Name :- Digital Multimeter	
Make/I.D No. :- Rish/PIS/DMM/02	
Certificate No. :- TSC/21-22/18060-1	
Certified By :- Transcal	
Calibration Validity :- 14-Mar-23	

<b>4.Calibration Results</b> :-				
Calibration Points VAC	UUC Reading VAC	Standard Reading VAC	Error in VAC	Error in %
20	20	19.988	0.012	0.06
80	80	79.984	0.016	0.02
140	140	139.76	0.24	0.17
200	200	199.87	0.13	0.06
270	270	269.79	0.21	0.08

**5.Accuracy in Measurement** :- ± 0.17 %

**6.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

- Note:**
- 1) UUC stands for Unit Under Calibration.
  - 2) This certificate refers only to the particular item submitted for calibration
  - 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
  - 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By 		Approved By 
Calibration Engineer Mr.Rahul Pawar		Technical Manager Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>Bush Electromech and Engineering Pvt. Ltd.</b> K-257, MIDC, Waluj, Aurangabad 431 136.	Page No. :- 1 of 1
Ambient Temp. :- 25 °C	Date of Receipt :- 6-Oct-22
Relative Humidity :- 54 % RH	Service Request No. :- A2022/22
Location of calibration :- On Site	Certificate No. :- A2022/10/075
Condition of Item :- OK	Date of Calibration :- 6-Oct-22
	Next Calibration Due On :- 6-Oct-23
	Calibration method No. :- PISPL/ET/WI/04

### 2. Details of unit under Calibration (UUC):-

Name :- Insulation Tester (Megger)	Range :- 0 to 5000 MΩ
I.D No. :- INS/MEG/01	Specified Accuracy :- ± 2 %
Make :- ---	Location :- ---
Type :- Analog	

### 3.Details of Equipment Used for Calibration

Name :- Multifunction Calibration System with Current Coil
Make/I.D No. :- Fluke/PIS/UNICAL/03
Certificate No. :- NEC/22-23/SP/065-1
Certified By :- NEC
Calibration Validity :- 22-Sep-24

### 4.Calibration Results :-

Calibration Points MΩ	UUC Reading MΩ	Standard Reading MΩ	Error in MΩ	Error in %
1	1	1.00263	-0.00263	-0.26
5	5	5.0285	-0.0285	-0.57
20	20	20.0472	-0.0472	-0.24
100	100	100.158	-0.158	-0.16
200	200	200.197	-0.197	-0.10

5.Accuracy in Measurement :- ± 0.57 %

6.Remark :- The Unit Under Calibration Found to Perform Satisfactorily.

#### Note:

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By

*Rahul*

Calibration Engineer

Mr.Rahul Pawar



Approved By

*oie*

Technical Manager

Mrs.Vinita Ghogare



# PRECISION INSTRUMENTATION AND SERVICES PVT. LTD.

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21, Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011. Tel.: 0253 - 6145100

M. : 09860102524 • Web site : www.pisnashik.com • E-mail : customercare@pisnashik.com

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:- 1 of 1
<b>Bush Electromech and Engineering Pvt. Ltd.</b>	Date of Receipt	:- 6-Oct-22
K-89, MIDC, Waluj, Aurangabad 431 136.	Service Request No.	:- A2022/22
Ambient Temp. :- 24.8 °C	Certificate No.	:- A2022/10/076
Relative Humidity :- 52.5 % RH	Date of Calibration	:- 6-Oct-22
Location of calibration :- On Site	Next Calibration Due On	:- 6-Oct-23
Condition of Item :- OK	Calibration method No.	:- PISPL/ET/WI/03

<b>2. Details of unit under Calibration (UUC):-</b>			
Name	:- Three Phase Secondary Injection Kit	Range	:- 0 to 2 KVAC
I.D No.	:- QMPL-052092019		
Make	:- Agronics		

<b>3.Details of Equipment Used for Calibration</b>			
Name	:- Digital Multimeter		
Make/I.D No.	:- Rish/PIS/DMM/02		
Certificate No.	:- TSC/21-22/18060-1		
Certified By	:- Transcal		
Calibration Validity	:- 14-Mar-23		

### 4.Calibration Results :-

#### 1.SET AC/DC VOLTMETER VERIFICATION

	SET VOLTMETER READING (V)	EXTERNAL STANDARD VALUE (V)
	63	63.4
	110	110.5
	230	229.7

#### 2.SET AC CURRENT OUTPUT VERIFICATION

SET AMMETER READING (Amp)	R PHASE (Amp)	Y PHASE (Amp)	B PHASE (Amp)
2	2.2	2.3	2.1
4	4.1	4.2	4.1
10	10.3	10.5	10.3
15	15.4	15.7	15.2
30	30.5	30.6	30.3

**5.Remark** :- The Unit Under Calibration Found to Perform Satisfactorily.

**Note:**

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
- 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By  
*Rahul*

Calibration Engineer

Approved By  
*ccc*

Technical Manager