



# CERTIFICATE OF CALIBRATION

NAME OF CUSTOMER Global Marketing company

SERIAL NO. 1533

W. O. NO. 161707635

CERT. NO. 132

DATE: 03/11/2013

## COMPARATOR STAND

Instrument Details -

MODEL - CS-3A

BASE - plain base

IDENTIFICATION NO. - 2014/50

Sr. No.	PARAMETERS	SPECIFIED VALUES	OBSERVED VALUES
1)	MEASURING RANGE	0-200	OK
2)	FINE ADJUSTMENT RANGE	-	-
3)	SQUARENESS OF COLUMN Wrt WORKING FACE	$\frac{0.030}{100}$	$\frac{0.020}{200}$
4)	PARALLELISM OF TABLE TOP PLANE	0.100	0.015
5)	FLATNESS OF BASE SEBRATED / PLAIN / GRANITE	0.003	0.0025
)	-	-	-

NOTE: All Dimensions are in m.m.

ALL OTHER DIMENSIONS ARE CHECKED AND FOUND WITHIN SPECIFIED LIMIT.  
EQUIPMENTS USED IN THIS CALIBRATION ARE TRACEABLE TO N.P.L. / N.A.B.L.

SPECIAL INSTRUCTION :

Checked By

Rshat  
 Manager Q. A.  
 Unity Gauge & Tool Co. Pvt.  
 Bhosari, Pune - 26.